

Certificate



Certificate no.

TA 50421410 01

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park of the annex list of factories
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1- 50181772 001

Client Reference: 180401303/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation	: CFM130MXY	10
X stands for	: 120, 240, 360 or 480	1
Y stands for	: blank, -B or -C	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 1.8-0.75A	
Outputs	: see constructional dataform	
max. Ambient Temperature:	: 40°C	
Protection Class	: I or II	

Continued on page 2

This certificate is valid for maximum 5 years.

12

Appendix: 1

Licensed Test mark:




TÜV Rheinland Taiwan Ltd

Signature

Date of Issue

(day/mo/yr)

13/11/2018



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50421410 02

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park of the annex list of factories
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1- 50181772 001

Client Reference: 180401303/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 1

Continuation

- Remark:
1. The certified product was evaluated for compliance with the requirements for both Class I and Class II. The actual classification shall be determined for the end product assembly.
 2. The risk management requirements and usability of the standard were not addressed.
 3. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Date of Issue

(day/mo/yr)

13/11/2018



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50421410 03

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park of the annex list of factories
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1- 50181772 003

Client Reference: 210800303/VTC

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation: 1) CFM130M180Y
2) CFM130M190Y

Y stands for : blank, -B or -C

Outputs : 1) DC 18V/5.56A, 100W (w/o fan cooling)
DC 18V/7.2A, 130W (with fan cooling)
2) DC 19V/5.26A, 100W (w/o fan cooling)
DC 19V/6.8A, 130W (with fan cooling)

Remark: 1. External cooling fan was applied with minimum inward air-flow of 10 CFM toward Q101 and L4 with distance of 5 cm.

2. This certificate is valid for maximum 5 years.

Appendix: 1

Licensed Test mark:




IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419063206

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

28/09/2021

Certificate



Certificate no.

TA 50421410 04

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd.,
Fu Hsing Park Fu Hsing Hsiang,
Chang Hua Hsien 506
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1-KM 50181772 004

Client Reference: 230503803 / VTC

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Change

Test Requirement : see above

- Remark:
1. The risk management and usability requirements of the standard were not addressed.
 2. The certified product was not classified. The actual classification shall be determined for the end product assembly.
 3. This certificate is valid for maximum 5 years.

Appendix: 2

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker

Date of Issue
(day/mo/yr)

12/10/2023

