

## UL TEST REPORT AND PROCEDURE

<b>Standard:</b>	<b>ANSI/AAMI ES60601-1:2005 &amp; A1:2012 &amp; A2:2021 (consolidated version) – revision date: 2022/02/01</b> <b>CAN/CSA-C22.2 No. 60601-1:14 (IEC 60601-1:2005+A1:2012+A2:2020, MOD) (consolidated version) – revision date: 2022/July</b>
<b>Certification Type:</b>	Component Recognition
<b>CCN:</b>	QQHM2, QQHM8 (Power Supplies, Medical and Dental)
<b>Product:</b>	Switching Power Supply
<b>Model:</b>	CFM21SX-Y (X can be 033, 050, 090, 120, 150, 240; -Y can be blank, -E, -T or -S)
<b>Rating:</b>	Input: 100-240Vac, 0.5-0.3A, 47-63 Hz  Output: - For models CFM21S033, CFM21S033-E, CFM21S033-S and CFM21S033-T 3.3 Vdc, 4.0A or 2.0A (half load)  - For models CFM21S050, CFM21S050-E, CFM21S050-S and CFM21S050-T 5.0 Vdc, 4.0A or 2.0A (half load)  - For models CFM21S090, CFM21S090-E, CFM21S090-S and CFM21S090-T 9.0 Vdc, 2.3A or 1.15A (half load)  - For models CFM21S120, CFM21S120-E, CFM21S120-S and CFM21S120-T 12 Vdc, 1.7A or 0.85A (half load)  - For models CFM21S150, CFM21S150-E, CFM21S150-S and CFM21S150-T 15 Vdc, 1.4A or 0.7A (half load)  For models CFM21S240, CFM21S240-E, CFM21S240-S and CFM21S240-T 24 Vdc, 0.9A or 0.45A (half load)
<b>Applicant Name and Address:</b>	CINCON ELECTRONICS CO LTD 8-1 FU KUNG RD FU HSING PARK FU HSING HSIANG CHANGHUA HSIEN 506 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.