

Certificate



Certificate no.

TA 50450356 01

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park of the annex list of factories
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Manufacturing Plant:

Refer to latest revision

Test report no.:

ZTW1- 50291980 004

Client Reference:

190500203/1905002047

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product:

(Switching Power Adapter)

License Fee - Units

Type Designation	: TR18RDMx	8
x stands for	: 050, 090, 120, 150, 180 or 240	1
Rated Voltage	: AC 100-240V, 50-60Hz	
Rated Current	: 0.5-0.25A	
Outputs	: see constructional dataform	
max. Ambient Temperature:	40°C	
Protection Class	: II	

This certificate is valid for maximum 5 years.

9

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Date of Issue

(day/mo/yr)

28/10/2019



Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50450356 02

License Holder:

Cincon Electronics Co., Ltd.
No. 8-1, Fu Kung Rd., Fu Hsing Park of the annex list of factories
Fu Hsing Hsiang, Chang Hua Hsien
506
Taiwan

Manufacturing Plant:

Refer to latest revision

Test report no.: ZTW1-KM 50291980 007

Client Reference: VTC/230504103

Tested to: IEC 60601-1:2005+A1+A2
EN 60601-1:2006+A1+A12+A2
IEC 60601-1-11:2015+A1
EN 60601-1-11:2015+A1

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Change

Test Requirement : see above

Remark: 1. The risk management and usability requirements of
of the standard were not addressed.
2. This certificate is valid for maximum 5 years.

Appendix: 2

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker

Date of Issue
(day/mo/yr)

06/09/2023

